



170021131268



中国认可
检测
TESTING
CNAS L0738



CONFORMITY CERTIFICATE OF EXPLOSION-PROOF

Certificate No.: CCRI 18.1228

Company: Shanghai Rocksensor Automation Co.,Ltd

Address: No.118 Lane315 Yuanqu Bei Road, North 2/F Bldg 3, Minhang, Shanghai

Name of Product: Multiparameter flow transmitter

Type of Product: RP1005

Marking: Ex d II C T6 Gb

Technical Documents: Q/LDS 001-2018

Drawing No.: RP1005-000

Note(s): 1. Rated voltage: DC 24V, Output signal: 4~20mA;
2. IP code: IP66.

By verifying the drawings and technical documents and checking samples, the product complies with the following standards currently valid in P.R.China:

GB 3836.1-2010, GB 3836.2-2010, GB/T 4208-2017

Date of Issue: 9. 26. 2018

Valid till: 9. 25. 2023

Authorize:



China Coal Research Institute Testing Center

(The test center of the national safety production of mining equipment and mining oil in Beijing)

Note: This certificate is only valid for the products that are in accord with samples tested and verified.
Center Add: No.5 Qingniangou Road, Hepingli, Beijing, 100013 Tel/fax: 010-84262992
The query url of Explosion-proof Certificate: <http://www.ccritc.com.cn/Home>